



E-beam Technology Industrial Deployment- Regulatory Framework and Quality Control Measures for Food Irradiation and Medical Device Sterilization

V.C. Petwal,
Industrial Accelerators Division, RRCAT, Indore
vikash@rrcat.gov.in




“Committed to take DAE technologies to commons for wellbeing”

Outline

- ✓ **Indian E- beam technology (10 MeV Linac based) and its industrial deployment**
- ✓ **Overview of E-beam facility and E-beam processing**
- ✓ **Regulatory framework for E-beam processing (food & medical devices)**
- ✓ **Licensing requirements and procedures**
 - ❖ **License for safe operation**
 - AERB “Type Approval” qualification (IQ, OQ)*
 - AERB License for operation*
 - ❖ **License for commercial irradiation**
 - Quality management system (QMS) : MDR-2017*
 - ISO 9001 & ISO 13485*
 - CDSCO/SLA license for MD sterilization*
- ✓ **E-beam sterilization & quality control: ISO 11137**
 - Process qualification and validation**
 - Product/ process specification**
 - Process control and integrity**
 - Product release and customer feedback**
- ✓ **Conclusion**


E-beam Technology: Emerging radiation processing service in India



Industrial deployment of RRCAT E-linacs



Linac-1 and Linac-2 (10 MeV, 6 kW) deployed at ARP



Endurance testing of Linac-3 at 10 MeV, 10 kW RRCAT




10 MeV, 10 kW Linac deployed at M/s MSSPL Bangalore



Linac-4 on test bench at RRCAT, 10 MeV, 15 kW

E-Beam Facility at RRCAT, Indore

The first E-beam facility (ARPF) in the country: providing commercial sterilization service to regulated medical devices



Designed , developed and operated by RRCAT


Salient features



Radiation source	10 MeV, 6 kW RRCAT electron Linac
Regulatory licenses	Plant operation: AERB, MD sterilization: CDSCO-FDA
QMS	MDR 2017
QMS accreditation	ISO 9001 and ISO 13485 for e-beam sterilization as per ISO 11137-1

QMS surveillance audit	CDSCO Notified body CE auditing
Processing capacity	200 kg per hour (for a minimum dose of 25 kGy)
Box dimension	590 x 430 x 340 mm (Typical)
Beam availability	100%


E-beam Technology





Product density (g/cc)	Single side	Double side
1	3.5 cm	8 cm
0.5	7 cm	16 cm
0.4	9 cm	20 cm
0.3	12 cm	26 cm
0.2	18 cm	42 cm
0.15	23 cm	55 cm

Useful penetration depth

Single sided Irradiation



Double sided Irradiation



Regulatory Framework for radiation processing of food and medical products

- ✓ Regulation ensures the safety, reliability, and quality of the process/service
- ✓ Regulation together with Quality Control ensures that product/service comply the requirement of standards and thus create public acceptance and trust.

Regulatory framework for radiation processing of food

1. Radiation processing of food items under Atomic Energy (Radiation Processing of Food & Allied Products) Rules, 2012 - Class wise
2. Food Safety and Standards (Packaging and Labeling) Amendment Regulations, 2016
3. AERB Safety Code No. AERB/RF-RPF/SC-1 (Rev.1) Radiation Processing Facilities

Salient features

Type of the radiation sources permitted

- a) Gamma radiation;
- b) Electron beams from a machine operated at or below ten million electron volts;
- c) [#]X-rays generated from a machine operated at or below seven and half million electron volts;
with *Tantalum or Gold as X-ray target material (USFDA)*





Plant operation

By competent qualified personnel for operation , quality control and safety

Training as recognizes by Atomic Energy regulatory board

Labeling

Mandatory to label the product processed by radiation with RADURA symbol



Classes of Food Products and Dose Limits: GoI Atomic Energy Rules 2012

Class	Food	Purpose	Dose Limit (kGray)		Class	Food	Purpose	Dose Limit (kGray)	
			Min.	Max.				Min.	Max.
Class 1	Bulbs, stem and root tubers and rhizomes	Inhibit sprouting	0.02	0.2	Class 5	Meat and meat products including poultry (fresh and frozen) and eggs	Elimination of pathogenic microorganisms	1.0	7.0
Class 2	Fresh fruits and vegetables (other than Class I)	Delay ripening	0.2	1.0			Shelf-life extension	1.0	3.0
		Insect disinfestation	0.2	1.0			Control of human parasites	0.3	2.0
		Shelf-life extension	1.0	2.5	Class 6	Dry vegetables, seasonings, spices, condiments, dry herbs and their products, tea and plant products	Microbial decontamination	6.0	14.0
		Quarantine application	0.1	1.0			Insect disinfestation	0.3	1.0
Class 3	Cereal and their milled products, pulses and their milled products, nuts, oil seeds, dried fruits and their products	Insect disinfestation	0.25	1.0	Class 7	Dried foods of animal origin and their products	Insect disinfestation	0.3	1.0
		Reduction of microbial load	1.5	5.0			Control of molds	1.0	3.0
Class 4	Fish, aquaculture, seafood and their products (fresh or frozen) and crustaceans	Elimination of pathogenic microorganisms	1.0	7.0			Elimination of pathogenic microorganisms	2.0	7.0
		Shelf-life extension	1.0	3.0	Class 8	Ethnic foods, military rations, space foods, ready-to-eat, ready-to-cook/minimally processed foods	Quarantine application	0.25	1
		Control of human parasites	0.3	2.0			Reduction of microorganisms	2	10
							Sterilization	5	25

Regulatory Framework for E-beam Sterilization of Medical Devices

1. ISO-11137: 2010

Sterilization of health care products Radiation — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices.

2. Medical Devices Rules (MDR-2017) GoI

3. Indian Pharmacopeia, Volume-1, 2018

4. ISO 13485: 2016

Medical devices — Quality management systems — Requirements for regulatory purposes




5. AERB Guide No. AERB/SG/IS-5 and AERB/RF/SG/G-3



IAEA RTS No. 4

Guidelines for the Development, Validation and Routine Control of Industrial radiation processing


IAEA Report


Trends in Radiation sterilization of health care products





Sterilization Process Qualification, Validation and Quality Control





Sterilization process development (PQ dosimetry)

Dosimetry Box No	D _{min} Dose (kGy)	D _{max} Dose (kGy)	D _{ref} Dose (kGy)
PQ-1	25.8	34.0	35.0
PQ-2	26.6	34.3	35.1
PQ-3	29.4	40.4	36.4
PQ-4	26.2	33.3	34.6
PQ-5	26.4	33.9	34.6
PQ-6	26.4	34.3	36.0
PQ-7	27.4	35.0	34.7
PQ-8	27.1	37.4	34.9
Average dose	<u>26.9</u>	<u>35.3</u>	<u>35.2</u>
S.D. ±	1.1	2.4	0.7
C. V. (%)	4.2	6.8	1.9

- ✓ Identify D_{min} and D_{max} position
- ✓ Establish relation between reference position dose and D_{min} & D_{max} dose
- ✓ Set operating parameters to deliver the required dose with 95% confidence level

Sterilization process validation



ACCUPREC
RESEARCH LABS PVT. LTD.
Ultimate solutions...
FDEA LICENSE NO. ST/17/3
Gurgaon, Haryana, India. 122001-122002, India

**RESEARCH
REDEFINE**

CERTIFICATE OF ANALYSIS Form: ST (Date: 05/01/2024) - Report of test or analysis by approved institution Product Product			
Product Name: Patel Dish		Lot No: 12/02/2024	
Party Name and Address: ARL India Pvt. Ltd., A-5 Malviya Nagar, Phase II, New Delhi		Expiry Date: 29/02/2024 to 22/02/2024	
Manufacturer License No: N/A	Doc No: N/A	Doc Date: 29/02/2024	Doc Rev: 01
Batch No: N/A	Doc No: N/A	Doc Date: 29/02/2024	Doc Rev: 01
Std. No: N/A	QMS No: N/A	QMS Date: 29/02/2024	QMS Rev: 01
Exp. Date: N/A	Exp. No: N/A	Exp. Date: 29/02/2024	Exp. Rev: 01
State of Sample Storage: 20/02/2024	EM No: N/A	EM Date: 29/02/2024	EM Rev: 01
Sample Id: ARL/1496/009-2024	Refuse Date: 22/02/2024		

No.	Item	Test Method	Standards & Limits	Observation
1	sterility test	USP 41	This test is used to determine the sterility of the product. The product should be incubated for 14 days of incubation. There should be no growth observed. The product should be incubated for 14 days of incubation. There should be no growth observed. The product should be incubated for 14 days of incubation. There should be no growth observed.	The growth test results to comply up to 14 days of incubation. There is no bacterial growth in 14 days of incubation. The product is sterile. The product is sterile. The product is sterile.

I, the undersigned, in the capacity of the undersigned, the sample referred to above is of standard quality and complies with the attached standards of USP 41.

Authorised By: **NAKASHA** Authorised By: **NAKASHA**
Date: 29/02/2024 Date: 29/02/2024

Standard5
Commitment to Excellence

TEST REPORT

Test Report No: 20240717/0003-71 Date: 27-07-2024 Page: 1 of 1

CUSTOMER DETAILS	
Customer Name & Address: M/A Princes Gloves (P) Ltd, Old Industrial Estate, Karamnagar, Jaipur	Test Requested: 12/02/2024
Product Group: Medical Accessories	Sample Received on: 17-07-2024
Manufacturer License No: N/A	Test Completed on: 17-07-2024
Batch No: N/A	Sample Collected on: 14-07-2024
Std. No: N/A	Sample Quantity & Packing: 1 500g x 5 Set
Exp. Date: N/A	Sample Quantity & Packing: 1 500g x 5 Set

INFORMATION AVAILABLE ON THE TEST STRIP	
Product ID: Princes Gloves	Exp. Date: 01/07/2024
Lot No: 007-19	Exp. Date: 01/07/2024
Date of Manufacture: 04/07/2024	Batch of Manufacture: 00032, ARFF/India

TEST RESULTS - BIOLOGICAL PARAMETERS			
Sl. No.	Sample Code	Sl. Position	Result
1	20240717/0003	01-01	No growth observed. Sterilizability cycle is satisfactory.
2	20240717/0003	01-02	No growth observed. Sterilizability cycle is satisfactory.
3	20240717/0003	01-03	No growth observed. Sterilizability cycle is satisfactory.
4	20240717/0003	01-04	No growth observed. Sterilizability cycle is satisfactory.
5	20240717/0003	01-05	No growth observed. Sterilizability cycle is satisfactory.

Reference incubation period is 7 days. (Index of Report)

Biology Management
Microbiology

Biology Management
Microbiology


Biology Management
Microbiology

Standard5 Environmental & Analytical Laboratories
Approved & Recognized by: **CCRI** (Central Council for Irradiation Dosimetry), **IAEA** (International Atomic Energy Agency), **ISO 9001:2015** (Quality Management System), **ISO 14001:2015** (Environmental Management System), **ISO 45001:2018** (Occupational Health and Safety Management System).

Third Party Sterility Testing at NABL Certified Labs


Quality Assurance at ARPF: Regular operation

QA measures to ensure that the irradiation process is in state of control to meet customer requirement





Process Control: Online Electronic Monitoring of Irradiator System

Monitoring of critical process parameters of 100 Batches processed Indicates that the irradiation process is in state of control.



Process Control: Dosimetry


Dosimetry data monitoring after processing of each batch of medical devices indicates that the Irradiation Process is in the state of control.



Radiochromic Film Dosimetry System




EPR Dosimetry System



Dosimetry data of 100 batches processed

Process Control: Customer product specification

- ❖ Customer registration, order registration
- ❖ Generation of unique dose mapping number for a product
- ❖ QC checks: for incoming material, Batch preparation
- ❖ Product tracking during irradiation and after irradiation
- ❖ Generation of unique bar-code for each irradiated box.
- ❖ Product processing IN and OUT time logging.
- ❖ Dosimetry data recording and logging
- ❖ Generation of "Product Irradiation Certificate" and product release related documents



AGRICULTURAL RADIATION PROCESSING FACILITY (ARPF)	
RAJA RAMANNA CENTRE FOR ADVANCED TECHNOLOGY DEPARTMENT OF ATOMIC ENERGY, GOVERNMENT OF INDIA, Indore- Madhya Pradesh	
Product Irradiation Certificate Date - 14/11/2024	
Certificate No: ARPF/PC/0009909	
ARPF/PC/0038/0038	
Product Details	
1. Client registration no.	: ARPF-REG-0007
2. Name	: M/s Becton Dickinson India Pvt. Ltd., Rewari, Haryana-123401
3. Product Name	: IV Canula
4. Product Batch No	: 16G-4291316, 20G-4291308, 4281543, 22G-4281270, 4290973
5. Quantity	: 472 cartons
6. PDM Reference Number	: ARPF-PDM-MD-568
Irradiation Process Details	
7. E-Beam Process Order No	: ARPF-Order-0029
8. Process Run Start Date	: 08-11-2024
9. Process Run End Date	: 13-11-2024
10. Minimum Specified Dose (kGy)	: 25 Average Minimum Delivered Dose (kGy) : 25.5
11. Maximum Specified Dose (kGy)	: 55 Average Maximum Delivered Dose (kGy) : 48.6
Also non-conformity reported during irradiation : No	
<div style="display: flex; justify-content: space-around; margin-top: 10px;"> <div style="text-align: center;"> QC In-charge </div> <div style="text-align: center;"> Radiation Processing In-charge </div> <div style="text-align: center;"> Facility In-charge </div> </div>	
ARPF provides electron beam irradiation services as per the QMS implemented to meet the requirements of ISO 13485 and in alignment with ISO11137. The product has been irradiated with the dose limits prescribed by client in the order registration form. ARPF certifies that the processed items, received the indicated dose within the precision and accuracy of the dosimetry system used. Address: Agricultural Radiation Processing Facility (ARPF), Indore in Tirthi Market, Dist: Alakhnagar, Madhya Pradesh and Vegetable Market Complex, Vidyanagar, Indore (M.P.) ARPF-OPR-10, Ver-01, Rev-01, dated 13-08-2021	

Certificate of Irradiation to release product

Irradiation Agreement for commercial sterilization service

Becton Dickinson, Haryana

Irradiation Agreement

Between
Becton Dickinson India Private Limited
and
Raja Ramanna Centre for Advanced Technology (RRCAT)

July 2022

IV Cannula

Product box labelling

Manufacturing Site & Customer Complaint Address: Becton Dickinson India Pvt. Ltd. Plot No. 1, Sector-1, IIT, Bawal, Rewari, Haryana-123401, INDIA, M.L. MFG/MD/2019/0000314
 Sterilization by Becton Dickinson India Pvt. Ltd. at Raja Ramanna Centre for Advanced Technology, Department of Atomic Energy, Government of India, Agricultural Radiation Processing Facility, Dist: Alakhnagar, Madhya Pradesh and Vegetable Market Complex, Indore, Madhya Pradesh, India, under license no. ARPF/0037/002925

Triveni Polymers, Haryana

Irradiation Agreement

Between
Triveni Polymers Pvt. Ltd.
and
Raja Ramanna Centre for Advanced Technology (RRCAT)

September 2022

PP vials



Agricultural Radiation Processing Facility, Indore

A CSCCO FDA, ISO 9001:2015 and ISO 13485:2016 Certified Facility

ARPF-10-2022-007-008-0003-0029-00

Customer: **Becton Dickinson India Pvt. Ltd.**

Registration Number: **ARPF-REG-0007**

Order Number: **ARPF-ORDER-0003**


Date of Irradiation: **08-10-2022**

Purpose of Irradiation: **Sterilization**

CONTENTS OF THIS PACKAGE HAVE BEEN IRRADIATED WITH ELECTRON BEAM AT ARPF, INDORE



Medical device sterilization (Progressive Growth)



Conclusion

- ✓ RRCAT has developed 10 MeV E-Beam Linac technology for industrial applications
- ✓ 10 MeV, 10 kW Linac technology is “Type Approved” by AERB.
- ✓ Three Linacs are operational in industry and providing e-beam irradiation service.
- ✓ The facility at Indore is operating under the regulatory framework of AERB and FDA and providing sterilization services for medical devices
- ✓ Implemented QMS comply the requirements of MDR-2017 and has accreditation for ISO 9001:2015 and ISO 13485:2016
- ✓ The facility is being used by medical device manufacturers from across the country for sterilization of their products on regular basis.



Thank You

“Electrons in the Service of Nation”



